

Attachment 1 Specifications.

1.5 LEASE SPECIFICATIONS

- A. Lease of Blood Pressure Kiosks will be in the form of a Services Contract and will initially be for a One (1) year term.
- B. Pricing will be monthly per Kiosk; delivered, set-up and fully functioning. Additional fees will not be charged or paid.
- C. A single invoice will be received monthly.
- D. At end of initial contract term upon mutual agreement between ADH and vendor the agreement may be renewed.
- E. At the final contract end or if ADH chooses not to renew the contract in accordance with the contract terms a 90-day notification will be issued to the vendor. All final 90-days invoices will not be paid by ADH until each unit is collected by the vendor per contract terms.

1.6 SPECIFICATIONS FOR BLOOD PRESSURE KIOSK STATIONS

- A. BP kiosk with monitor, seat, and BP cuff all in one compact kiosk.
- B. Complies with the ANSI/AMMI/ISO standard for blood pressure measurement accuracy.
- C. Oscillating method for stepwise pressure deflation technique.
- D. Test time 1 minute or less.
- E. Easy to use, one button operation.
- F. Compact size no larger than 30 inches x 24 inches floorspace.
- G. Height 63 inches or less.
- H. Weight 150 lbs. or less.
- I. Color monitor with no less than 6-inch 640x480 resolution.
- J. The monitor displays systolic and diastolic pressures and pulse rate.
- K. Over pressure limit at least 290 mmHg minimum.
- L. Measurement range no less than: Systolic 60-255 mmHg; Diastolic 25-220mg Hg.
- M. Cuff size 20cm-43cm.
- N. Input power is standard 3-wire grounded outlet 100-120V.
- O. Operating temperatures No less than 50 -102° F.
- P. Print function not required.
- Q. Internet connection capability not required.
- R. Ability for on-site ADH personnel to obtain usage reports. Vendor will provide instructions.
- S. Used/refurbished equipment meeting all specifications is acceptable.

1.7 SERVICE AND DELIVERY:

- A. Please refer to Attachment 2 LHU Blood Pressure Kiosk Project List.
- B. Delivery will be to each Local Health Unit by the vendor. ADH will not provide a central shipping receiving location.
- C. Delivery and Set-Up will commence in 2 phases.
 - 1. Phase 1 will place Blood Pressure Kiosks at 22 priority clinics highlighted on Attachment 2.
 - 2. Phase 1 On-site delivery, set-up, operator instructions, and full functioning of all 22 units will be completed by March 30, 2024.
 - 3. Phase 2 (the remaining 67 LHUs) will commence after completion of Phase 1.
 - 4. Phase 2 will be completed no later November 30, 2024.

5. Alternative schedules proposed by bidders shall be detailed on the Exceptions sheet in the Bid Response Packet. If a responsive bid is not received that meets the planned schedule ADH may accept an alternate schedule from the lowest priced otherwise responsive bidder.

- D. Vendor will provide a minimum 2 week notice for delivery to each individual site contract manager and site-specific contact person (contact list will be provided with contract)
- E. Delivery & Set-up of each unit includes removal and proper off-site disposal of all packing material, fully set-up and fully tested functioning and first invoice submitted. First month invoice to be pro-rated if delivery and Set-up completed after the 15th day of a Month.
- F. Invoices will be submitted monthly for all Kiosks delivered, set up and fully functioning. Kiosks in place but not functioning will not be invoiced or paid.
- G. A USA based Toll-free telephone number will be provided for service and support.